

Technical Product Report

For Research Use Only; Not for use in Diagnostic Procedures

Product Description: Seraseq™ ctDNA Complete Reference Material WT

Material No: 0710-0674

Batch No: 10346910

Material Description: Human genomic DNA from the reference cell line, GM24385

Date of Manufacture: 17May2018

Expiration Date: 17May2020

Concentration (Qubit dsDNA BR Assay): Nominal value: 25 ng/mL; Average measured value after extraction using Qiagen QIAamp Circulating Nucleic Acid Kit: 31.8 ng/mL

Volume: 5 mL

Storage: 4 °C

Digital PCR testing using
BioRad QX200™ Droplet
Digital™ PCR System:

Gene ID	COSMIC Identifier	Amino Acid Change	Average AF% ¹
AKT1	COSM33765	p.E17K	0.005
BRAF	COSM476	p.V600E	0.000
EGFR	COSM6224	p.L858R	0.001
EGFR	COSM6240	p.T790M	0.009
ERBB2	COSM682/20959	p.A775_G776insYVMA	0.007
KIT	COSM1314	p.D816V	0.002
KRAS	COSM521	p.G12D	0.012
NCOA4/RET	NA	Translocation	0.000
NRAS	COSM584	p.Q61R	0.006
PIK3CA	COSM775	p.H1047R	0.006
PIK3CA	COSM12464	p.N1068fs*4	0.006
EML4-ALK	NA	Translocation	0.001
ALK	COSM144250	p.G1202R	0.000
ALK	COSM28055	p.F1174L	0.000
BRCA1	COSM1383519	p.K654fs*47	0.001
BRCA2	COSM1738242	p.R2645fs*3	0.000
EGFR	COSM12370	p.L747_P753>S	0.000
EGFR	COSM6256	p.S752_I759delSPKANKEI	0.001
EGFR	COSM6223	p.E746_A750delELREA	0.008
KRAS	COSM516	p.G12C	0.001
CD74/ROS1	NA	Translocation	0.000
KRAS	COSM554	p.Q61H	0.004
Average AF%			0.003

NA = not applicable

¹Variant allele frequencies > 0.00% for this wild-type negative control are within the expected range for stochastic positive dPCR reactions.

NGS was performed as an orthogonal verification step. Results confirm no variants were detected above 0.1%, except for EGFR T790M variant detected at 0.13% in the WT sample; the corresponding dPCR data was 0.009%. We observed similar background in other Seraseq ctDNA Complete™ reference materials using this NGS assay, which were not observed in dPCR QC testing. Thus, we associate the allele frequency for the T790M variant with the background from the assay¹ using the conditions listed below.

Next Generation
Sequencing testing using
Archer® Reveal ctDNA™ 28
Kit run on an Illumina®
MiSeq™ using v2 (2x150 bp)
PE chemistry reagents:

NGS Parameters:

DNA input = 50 ng
of samples / flow cell = 3-4
of total reads / sample = 4-5M
Average read depth = 5000-10000X
On-target reads = ~94%
Q30 score = ~95%
Analysis = Archer Analysis Suite v5.1.7 (with error correction set to "ON")

¹Please see the poster from NIST for more information about assay sensitivity:

<https://digital.seracare.com/multilab-assessment-reference-materials-ctdna-poster2018>

Note: Copy numbers of ERBB2 and MET were assayed by dPCR and found to be normal. They were not assayed by NGS as the wild-type sample is used as a normal control for determining copy number of genes in other samples.

Approval:

A handwritten signature in black ink, appearing to be "J. R. A.", written over a horizontal line.

Prepared By

03/21/19

Date